

Biosensors: Classification, Fundamental Characterization and New Trends: A Review

Hardeep Kaur, Aakash Bhosale, Suraj Shrivastav

Department of Biotechnology, Chandigarh University, Gharuan Mohali, Punjab, India

Corresponding Author: Hardeep Kaur

ABSTRACT

Recent advances in developing low-cost and highly efficient biosensors devices which are highly sensitive and specific have opened new scope for detection of sample e.g. drugs, contaminated bio agents, metabolites, pollutants in atmosphere, and microbial load etc. by converting biochemical signals into measurable physiochemical signals which lead to quantify the amount of sample. Recently nano-biosensors, implanted biosensors and integrated biosensors are playing a critical role in current research area and development techniques. The various types of biosensors such as enzyme-based, tissue-based, immunosensors, and DNA biosensors, thermal and piezoelectric biosensors have been studied here to highlight their consequential applications in countless fields. Basic research is still required to refine the sensing strategies as well as analytical instrumentations and procedures to get new applications in numerous fields. Biosensors and their role in medical science including early stage detection of various diseases include human interleukin-10 causing heart diseases, rapid detection of human papilloma virus, etc. are critical aspects. Fluorescent biosensors play an important role in drug discovery and in cancer. DNA biosensors, based on nucleic acid recognition methods, are being developed towards the assay of rapid, simple and economical testing of genetic and infectious diseases. Moreover, the detection of specific DNA sequence is of importance in various areas of interest including clinical, environmental and food analysis. Biosensor applications are prevalent in the plant biology sector to find out the missing links required in metabolic processes. In this article, we present the basics of biosensing devices which can serve as an introductory part for those who are new to this field.

Keywords: Biosensors, sensing elements, transducer, amperometric, potentiometric, optical biosensor.

INTRODUCTION

Biosensors are hybrid of chemical and physical sensing technique which is one of the newly described classes of sensors. The definition of biosensors was given by IUPAC; according to him biosensors are chemical sensing analytical device that converts response into electrical signals which are highly specific independent of physical parameter like pH and temperature. [1-3] The term “biosensor” was given by Cammann. The biosensors are specifically based on receptor –transducer based tool,

which could be implemented for elucidate the biochemical and biophysical properties of the medium. Furthermore, biosensors have interesting character which makes them unique from others that they have biological/organic recognition element which helps to identify of specific biological molecules. Cells, antibodies, enzymes or nucleic acids can be considered as sensing elements which form s a recognition layer that is integrated with transducer via clogging by adsorption, covalent bonding or cross linking. [4] It is a versatile and

interdisciplinary technique which fascinates researchers regarding exploiting this field for other medical and future application by the collaborating efforts of physics, chemistry, engineering, biology and biotechnology and so on. [5] Biosensor gives great offer for several insider systematic

applications of quick and low cost measurement as it eradicates sample preparation. [6] Biosensor is preferred for food analysis, [7] bioterrorism, [8-9] environmental [8-10] and monitoring of human health like sweat wearable biosensor. [11-13]




Fig: 1. A scheme of Representation of biosensors.

Biosensor has been extensively used in contrastive scientific practice which can be due to their impressive outcome. According to the medical sciences, tumors, pathogens, inflated blood glucose levels in diabetic's patients and other toxins can be detected in definite and rigorous manner with the help of biosensors. For researchers biosensor which was encoded by genes generally referred as fluorescence generating biosensor have extreme relevance for examine and scrutinizing the complex chemical processes occurring inside the cells and these can be used to point out some specialized site present in the cell and could be expressed in particular cell of an organism. The whole deal solidification of a specific substance into the host cells could in like manner be expert through these biosensors. [14] On the off chance that there ought to be an event of food industry, biosensors could be associated with the recognizable proof of gasses released from destroyed support, revelation of sustenance contamination or for checking and constraining the improvement of tiny life forms or then again development in new sustenance. [15] From natural perspective, these biosensors could

be improved to distinguish contamination in air and nearness of any pathogens, overwhelming metals and so on. [16] In military protection frameworks, they can be utilized to recognize the nearness of any unsafe organic materials that would somehow or another stay imperceptible and cause demise. For this situation for the most part the biosensors can be utilized to distinguish the bioterrorist assaults like the deliberate utilization of the organic substances like Bacillus anthracis, Ebola, hepatitis C infections and so on. [17]

Here in, endeavors have been put into giving the basic investigation of the latest progressions in biosensing innovation gained in the traverse of late years and contextual analyses with particular fields and very high reference numbers have been accumulated to create a flexible audit of this field.

Types of Biosensors

The pioneers named Clarke and Lyons began Biosensor in late 1960s. Different kinds of biosensors being utilized based on two elements namely known as sensing element and transduction modes. Enzymes based biosensor, immunosensor which includes antibodies, DNA biosensor,

Thermal and piezoelectric biosensor, biological tissues, organelles and microorganisms which can be detected with the help of whole cell biosensor comes under the category of sensing element. Transduction mode relies upon the physiochemical change coming about because of detecting component. Subsequently on the premise of various transducers biosensors can be electrochemical (amperometric, conductometric and potentiometric), optical (absorbance, fluorescence and chemiluminense), piezoelectric (acoustic and ultrasonic) what's more, calorimetric. [18] Biosensors can likewise be arranged in view of their revelation arrange into original which is the easiest approach including direct discovery of either increment of an enzymatically created item or lessening of a

substrate of a redox chemicals utilizing characteristic go between for electron exchange e.g. glucose biosensor which utilizes chemical glucose oxidase and oxygen recognizing diminish in oxygen level or increment in hydrogen peroxide relating to the level of glucose. Second era biosensors utilizes manufactured redox middle people like ferrocene, ferricyanide and quinones for electron exchange which builds the reproducibility and affectability e.g. self-observing amperometric glucose biosensors. In conclusion, third era in which the redox compounds are immobilized on the cathode surface in such a way, to the point that immediate electron exchange is conceivable between the compound and transducer. It utilizes natural directing material e.g. TTF-TCQN (tetrathiafulvalene tetracyanoquinodimethane). [19]




Fig: 2. Types of biosensors applied in various fields of engineering.

The principal catalyst based sensor was accounted for by Updike and Hicks in 1967. Catalyst biosensors have been conceived on immobilization techniques, i.e. adsorption of proteins by van der Waals powers, ionic holding or covalent holding. The usually utilized proteins for this reason for existing are oxidoreductases, polyphenol oxidases, peroxidases, and aminooxidases. [20-22] The primary microorganism based or cell-based sensor was realized by Diviès. [23]

The tissues for tissue-based sensors emerge from plant and creature sources. The analyte of intrigue can be an inhibitor or a substrate of these procedures. Rechnitz [24] created the main tissue based sensor for the assurance of amino corrosive arginine. Organelle-based sensors were made utilizing layers, chloroplasts, mitochondria, and microsomes. Be that as it may, for this sort of biosensor, the steadiness was high, yet the recognition time was longer, and the

specificity was diminished. The primary sorts are examined beneath:

1. Amperometric biosensor

Electroactive species present in natural test samples can be easily detected by high affectability biosensor. The oxidation or diminishment of electroactive species is estimated and connected to the centralization of the analyte e.g. glucose biosensors for diabetes checking which produces current due to the potential difference between two electrodes. These anodes limit Eventually Tom's scrutinizing the preparing of a current at plausibility might be associated between two cathodes, the degree from guaranteeing current constantly corresponding of the substrate centralization. The Clarke oxygen cathode for presence of oxygen in the test (analyte) result during reduction is used by these biosensors for facing less difficulty. A main problem of such biosensors is their dependence on the separated O₂ fixation in the analyte result. This may be beat Eventually Tom's examining using go between; these particles trade the electrons made Toward the reaction clearly to the cathode as opposed to diminishing those O₂ deteriorated on analyte result. Those present-day anodes, in any case, remove those electrons particularly beginning with the diminished proteins without the assistance of arbiters, moreover require help secured with electrically coordinating normal salts. So, the natural test samples may not be characteristically electro-dynamic, catalysts are expected to catalyze the generation of radio-dynamic species. For this situation, the deliberate parameter is present. [25]

2. Potentiometric biosensor

In this strategy the scientific data is acquired by changing over the biorecognition procedure into a potential flag which results in oxidation or decreasing capability of biochemical responses. A perm selective particle conductive layer is typically used to gauge the potential flag, which happens when the analyte atom collaborates with the surface e.g. utilization

of H⁺ particles for penicillin discovery utilizing chemical penicillinase, triacyl glycerol utilizing lipase. A high impedance voltmeter is used to check the electrical potential qualification or electromotive power (EMF) between two cathodes. One of the terminals builds up an adjustment in potential as a component of analyte movement or fixation in arrangement and this cathode is known as the pointer anode or now and again called a ion selective electrode (ISE). The potential reaction of an ISE is portrayed by the Nernst condition (i.e., the potential is relative to the logarithm of the centralization of the substance being estimated). The second cathode is the reference and is utilized to finish the electrochemical cell by giving a consistent half-cell potential, which is autonomous of the analyte fixation. ISEs are compound sensors with the longest history and with the biggest number of uses. [26-28] Actually, billions of estimations are played out every year in almost every clinic everywhere throughout the world. [28] This shocks no one considering that these gadgets are outstanding for giving immediate, fast, upkeep free and non-costly estimations. [28-29]

3. Optical Biosensor

This type of biosensor can be based on the principle of optical diffraction or electro chemiluminescence in which a silicon wafer is covered with a protein by means of covalent bonds which is then presented to UV light through a photograph veil and the antibodies wind up idle in the uncovered areas. At the point when the diced wafer chips are brooded in an analyte, antigen-counter acting agent ties are shaped in the dynamic locales, consequently making a diffraction grinding. This grinding produces a diffraction flag when lit up with a light source. Optical biosensors comprise of a light source, and additionally various optical segments to produce a light bar with particular qualities and to shortcut this light to a balancing operator, an adjusted detecting head alongside a photo detector. [30] These biosensors measure both reactant

What's more regular slant reactions. They measure an advance for fluorescence on the other hand on absorbance expedited toward the outcomes created toward reactant reactions. On the other hand, they measure those movements provoked in the inborn optical properties of the biosensor surface in view of stacking on it for dielectric particles, for example, protein (in circumstance from asserting normal slant responses). A vast bit ensuring, biosensor coordinating, including brilliance usage firefly impetus luciferase for recognizable proof of infinitesimal life forms secured close by sustenance then again clinical tests. The minute life forms require help especially lysed should release ATP, which is used toward luciferase in the region around O_2 to plan light which is estimated Eventually Tom's scrutinizing the biosensor. [31]

4. Conductometric Biosensor

The deliberate parameter is the electrical conductance/protection of the arrangement. Conductometric-based biosensors saddle the connection amongst conductance and a biorecognition occasion. Most responses include an adjustment in the ionic species focus and this can prompt an adjustment in the arrangement electrical conductivity or current stream. [29] Basically, a conductometric biosensor comprises of two metal terminals (generally platinum or silver) isolated by a specific separation. Typically an AC (exchanging current) voltage is connected over the terminals, which makes a present stream be maintained between them. Amid a biorecognition occasion the ionic creation changes and an Ohmmeter (or multimeter) is utilized to quantify the adjustment in conductance between the metal cathodes. Some current investigations have demonstrated that this procedure is prepared to do quickly distinguishing (<10 mins) different sustenance borne pathogens (i.e., *Escherichia coli* O157:H7, *Salmonella*). [32-33] Alocilja and colleagues utilized a conductive polyaniline mark in the sandwich immunoassay plot, which

essentially enhanced the affectability by means of the development of a conductive sub-atomic scaffold between the two cathodes. [33-34] Sadly, one of the significant issues with this strategy is that the affectability is for the most part sub-par contrasted with other electrochemical techniques. [29]

5. Calorimetric Biosensor

Numerous catalysts catalyzed response are exothermic creating heat which is utilized as a reason for estimation of rate of response and consequently analyte focus. The temperature changes are resolved by thermistors e.g. cholesterol biosensors utilizing cholesterol oxidase (warm yield 53 $KJmol^{-1}$). The analyte game plan is experienced a little stuffed bed area containing immobilized substance; the temperature of the course of action is settled just before section of the game plan into the fragment and also as it is leaving the portion using separate thermistors. This will be those in every way that really matters everything thought about fitting kind of biosensor, utilizing no less than two proteins of the pathway in the biosensor on join two or three responses with broaden those glow yield. On the other hand, multifunctional proteins may a chance to be utilized. An example is the utilization of glucose oxidase for confirmation about glucose. [35]

6. Piezoelectric biosensor

Piezoelectricity can be clarified as a straight cooperation amongst mechanical and electrical frameworks in non-driven gem or comparative structure which initially find by Curie siblings in 1880. [36] Basically, the piezoelectric construct biosensor working in light of the vital that a swaying gem resounds at a characteristic reverberation recurrence. [37-38] The essential components in a biosensor are transducer and biorecognition component. Consequently, in piezoelectric biosensor the transducer is made of piezoelectric material (e.g., quartz) and the biosensing material that secured on the piezoelectric material which vibrate at the basic repeat. The recurrence is control by the outer electrical

flag which delivers a specific estimation of current, when the objective analyte is presented to the detecting material the connection/response will cause the recurrence move which will create changes in current perusing that can be examined to the mass of the analyte of intrigue. There are two essential sorts of piezoelectric sensors: mass wave (BW) and surface acoustic wave (SAW). In any case, writing indicates piezoelectric sensors are not get much consideration and second rate contrasted with electrochemical and optical based biosensing.

The bulk wave, quartz precious stone microbalance and surface acoustic wave transducer is on a very basic level in view of the piezoelectric impact. The extraordinary properties of piezoelectric material are used in this sort of detecting. Quartz is the most customarily used piezoelectric since it is unassuming, can be taken care of to yield single valuable stone and can withstand creation, warm and mechanical weight; regardless, there is report that lithium niobate and lithium tantalate can also be used. [38] A current survey has demonstrated that this strategy is exceptionally engaging when coordinate with Microelectromechanical frameworks (MEMS) for biosensing application. [39] Moreover the audit expresses that this kind of transduction is appropriate for touchy, compact and constant biosensing. [39] Piezoelectric transducer has been broadly connected and grasped for immunosensing application. [40-42] Some report recommends that the piezoelectric transducer is sensible for DNA and protein recognizable proof with revelation purpose of constraint of 1 ng/cm². [43] A few articles have show up in the writing detailing the utilization of piezoelectric sensor in different application, for example, cholera poison indicative location, hepatitis B, hepatitis C, sustenance borne pathogen recognition and so forth. [41,44-47] More importantly, it was revealed that piezoelectric is astoundingly sensitive strategy, seeing that a revelation limit of 8.6 pg/l was procured for hepatitis B disease

DNA and 25ng/mL for cholera harm area. [45-46] The favorable circumstances utilizing this sort of transduction are the continuous checking, mark free recognition and straightforwardness of utilization. [46, 48] Be that as it may, there are a few downsides need to defeat, for example, specificity, affectability and in addition obstruction decrease. [49] Likewise, this kind of transducer strategy includes organization and alignment necessity. [50] A current audit by Kim et al. 2011 has survey the guideline and use of nano symptomatic for nanobiosensor. The survey likewise finished up, that application scope of the quartz precious stone has been bit by bit extended, new estimating systems that utilization the quartz gem as a transducer for synthetic sensors and biosensors has been additionally created. [51]

7. DNA Biosensor

The classification of biosensors utilized for DNA discovery is otherwise called biodetectors. The objective is to disengage and measure the nature of single DNA– DNA or immune response antigen bonds, which along these lines has any kind of effect in recognizing and depicting single particles of DNA or antigen. The use of nucleic acids progression for the specific diagnostics application has made since the mid-1953 and up 'til now growing by and large. [52] The astoundingly specific proclivity limiting's reaction between two single strand DNA (ssDNA) chains to shape twofold stranded DNA (dsDNA) is utilized as a part of nucleic acids based biosensor which designate the nucleic acids as common affirmation segment. This technique has advanced the improvement of DNA based sensor from the conventional technique, for example, coupling of electrophoretic detachments and radio isotropic which are high cost, perilous, tedious and so on. [53] This biosensor working foremost depends on acknowledgment of the corresponding strand by ssDNA to frame stable hydrogen bond between two nucleic acids to end up dsDNA. In request to accomplish this, an immobilized ssDNA

is utilized as test in bioreceptor which the base grouping is correlative to the objective of intrigue. Presentation of focus to the test which brings about hybridization of correlative ssDNA to shape dsDNA will bring about creating biochemical response that permits transducer opened up the flag into electrical one. In this manner, writing demonstrates that the present of some linker, for example, thiol or biotin is required in the push to immobilize the ssDNA onto the detecting surface. [54] An imperative property of DNA is that the nucleic corrosive ligands can be denatured to invert authoritative and the recovered by controlling cushion particle fixation. [53] The nucleic corrosive natural acknowledgment layer which consolidates with transducer is effectively synthesizable, profoundly particular and reusable after warm dissolving o the DNA duplex. [55] Furthermore, this biosensor has an exceptional specificity to give expository instruments that can gauge the nearness of solitary particle animal categories in a complex blend. [56] DNA based biosensor has potential application in clinical indicative for infection and malady recognition. [57-59] In any case, electrochemical transduction is most desert technique used to examining DNA harm and connection which revealed in writing. The improvement of electrochemical DNA biosensor has gotten an incredible arrangement of consideration recently and this has generally been driven by the need to created fast reaction, high affectability, great selectivity and exploratory accommodation. [52] As an electrochemical gadgets are extremely helpful for succession particular bio-detecting of DNA. The scaling down of gadgets furthermore, propelled innovation make them brilliant device for DNA diagnostics. Recognizable proof of electrochemical DNA hybridization all the more frequently excludes checking a current at settled potential. Electrical modes were made for area of both name free and named objects. [60-71] The Fixation of the nucleic destructive test onto the surface of

transducer expects a focal part in the general execution of DNA biosensors and quality chips. [72-74] The immobilization step requires a specific depicted test introduction similarly, open to the objective transducer, diverse techniques can be utilized for connecting the DNA test to the strong surface, for example, the utilization of thiolated DNA test for self-gathered monolayers (SEM) onto gold transducers by covalent linkage to the gold surface through practical alkanethiol-based monolayers. The other strategy for connection of DNA test is to biotinylate DNA test and connection through biotin-avidin communication on terminal surface. [60-61,75-76] The avidin altered polyaniline electrochemically stored onto a Pt plate terminal for coordinate identification of E. coli by immobilizing a 5' biotin named test utilizing a differential heartbeat volta metric strategy in the nearness of methylene blue as a DNA hybridization marker. [60, 77] Similarly, electrochemical DNA biosensor in light of polypyrrole-polyvinyl sulfonate covered onto Pt plate terminal was additionally manufactured utilizing biotin-avidin restricting. [61] The revelation of carbon nanotubes (CNTs) in DNA examination plays a critical part by advancement of electrochemical DNA biosensor. CNT empowers immobilization of DNA atoms as well as utilized as effective speaker to enhance flag transduction of hybridization. CNT additionally fills in as novel pointer of hybridization. The use of exhibited CNT into DNA chip require little measure of test and advancement of CNT construct biosensor assume real part with respect to DNA based diagnostics in clinics or at home. [78] The learning of peptide nucleic corrosive (PNA) has opened another look into region of DNA biosensors. PNA is a DNA copy in which the sugar phosphate spine is supplanted with a pseudo peptide. The hybridization and affirmation of most essential solution phase PNA can be expeditiously extrapolated onto the surface of transducer for maintaining the relationship with the framework of

extraordinarily specific DNA biosensors. As such usage of surface kept PNA certification layers gives striking movement specificity onto DNA biosensors involving affirmation of single base confounds.^[74] The hybridization is usually distinguished by the expansion in current motion because of redox marker (that perceives the DNA duplex) or from other hybridization-instigated changes in electro synthetic parameters (e.g. conductivity or capacitance). New redox markers, offering more prominent segregation between single strand (ss) and dsDNA.^[79-80,62-63,65-66,81] The utilization of an entomb calator ferro ceryl naphthalene di-imide that ties to the DNA crossover more firmly than regular bury calators and shows little proclivity to the single-stranded test.^[82] The electrochemical DNA biosensor might be named based what's more, labeled free.

8. Surface Plasmon Biosensor

Surface plasmon reverberation (SPR) biosensor use surface plasmon waves (electromagnetic wave) to recognize changes when the target analyte team up with biorecognition segment on the sensor. On an essential level, when the SPR biosensor is displayed to any movements, it will incite changes in the refractive record which used to measure or viewed the reaction. The SPR transducer is uniting with biomolecule/ biorecognition segment which see and prepared to participate with specific analyte.^[83] Consequently when target analyte speak with the immobilized biomolecule on the sensor surface, it makes a modification in the refractive rundown at the sensor surface.^[83] This, movements convey an assortment in the spread unflinching of the surface plasmon wave and this assortment is measure to make examining. A spectrophotometer is used to check the maintenance scope of test. There been different biorecognition component have been fuses with SPR biosensor, for example, protein, counter acting agent antigen, nucleic acids and compound.^[84-88] A critical component of SPR biosensor is that it can give name free detecting without

radioactive and fluorescence which makes it exceedingly appealing for continuous checking.^[89-90] Moreover, the SPR based transduction can be utilized to and association without show any uncommon properties of fluorescence or trademark retention and diffusing groups.^[91] Notwithstanding, a few reports propose that these technique has experience the ill effects of specificity due to non-particular association with biorecognition component which wrongly corresponded by these biosensor.^[91] The SPR based transductions are not reasonable for examining little analytes. Since the mass of the material are measured by SPR related to the authority of the sensor's surface whereas small analytes ($M_r < 1000$) give little reactions.^[92] The current changes in flag to clamor proportion have made it conceivable to quantify official of such little analytes.^[92] SPR biosensors can adequately identify authoritative by atoms as little as around 2 kDa, yet littler particles create deficient changes in bound mass thus can't be specifically estimated satisfactorily.^[93] To date, SPR has been broadly utilized as a part of principal organic examinations, wellbeing science investigate, sedate find, clinical conclusion and ecological and horticulture checking.^[94] A few articles have showed up in the writing looking into the utilization of SPR based biosensor in pathogen and infection discovery.^[95-96]

9. Chemiluminescence based Biosensor

Chemiluminescence can be depicting as technique for vitality deliver from substance response which create an emanation of light or generally depict as Luminescence.^[97] Basically, when a concoction response happen, the particle or atom unwinds from energized state to its ground state which then a glow is create as side result of the response. Thusly, chemiluminescence can be utilized to see particular biochemical responses which happen besides, this property has contributed for chemiluminescence based biosensor progress. In the

chemiluminescence biosensor, the response among analyte and the immobilized biomolecule which has been separate with chemiluminescence species will end in conveying light as delayed consequence of biochemical response. This transmitted light can be recognized using a Photo Multiplier Tube (PMT). An audit by Dodeigne et al. 2000 and Zhang et al. 2005 has demonstrated that chemiluminescence is a developing apparatus for diagnostics with to a great degree high affectability along with the straightforward instrumentation, quick unique reaction properties, and wide adjustment go. Not with standings, chemiluminescence based transduction has been broadly connected and grasped for immunosensing and nucleic corrosive hybridization. [98-101] So also, various papers have demonstrated that chemiluminescence applications in clinical, pharmaceutical, natural and sustenance examination. [102-105] These techniques were likewise fused with immunosensor and optical fiber for discovery of dengue infection in human. [98] All the more imperatively, it was uncovered that this sort of transduction has recognition point of confinement of 5.5×10^{-13} M. [99] In any case, chemiluminescence transduction has couple of downsides, for example, less quantitative exactness because of short lifetime and not reasonable for ongoing observing. [106-108]

10. Fluorescence based Biosensor

The term Luminescence as portray above is the result of particles or atom which unwinds from energized state to its ground state. The different sorts of iridescence contrast from the wellspring of vitality to acquire the energized state. This vitality can be provided by electromagnetic radiation (photoluminescence likewise named as fluorescence or brightness), by warm (pyroluminescence), by frictional powers (triboluminescence), by electron affect (cathodoluminescence) or by crystallization (crystalloluminescence). [109] In this way, the fluorescence requires outside light source (short-wavelength light)

to begin the electronic advances in particles or ions which by then convey brilliance (Longer wavelength light). Over the long haul, the fluorescence-based biosensor has joined with fluorochrome particles which used to convey light in the midst of the biorecognition event. [110] Since an extensive part of the common identifying segment and most analyte does not have trademark frightful properties, the biorecognition event is transduced to the optical banner by coupling fluorescence optically responsive reagents to the recognizing segments. [111] For instance, the nucleic corrosive or antibodies are utilized to tag with fluorochrome and change over the hybridization collaboration between two correlative DNA remains into an optical flag. [112-114] Indeed, fluorescence innovation has been broadly connected and grasped for immunochemical detecting in the therapeutic field. [115-116] Fluorescence based biosensing were improved broadly in ecological observing as revealed by Védrine et al. 2003. In any case a current article has been show up in the writing on the utilization of fluorescence stamping DNA biochip for examination of DNA-cancer-causing agent adducts. [117] The significant downsides of fluorescence innovation are extra many-sided quality of time-settled instrumentation, in either the time or recurrence spaces or both and not appropriate for continuous checking. [118-119]

11. Immuno Biosensor

An antibodies based biosensor was connected out of the blue to recognition in the 1950s, opening the ways to the likelihood of immuno-conclusion. [120] From that point forward, there have been enthusiastic exertion made to create immunosensor which made out of antigen/immunizer as bioreceptor as a device for clinical diagnostics. [121-122] A counter acting agent is 'Y' formed immuno globin (Ig) that is comprised of two substantial chains (H) what's more, two light chains (L). However, some of human antibodies shape dimeric or pentameric structure by using disulphide bonds and an

additional protein called the joining or J-chain. [123-124] Every one of the chain has a steady and variable part. The variable part is particular to the antigen that is tie with relating antigen which is very particular furthermore, specific. [120-121] Thus, an immunosensor which made out of antigen as bioreceptor uses the capacity of immune response to tie with comparing antigen which is profoundly particular, stable, and adaptable. The specificity of an immune response towards the coupling side of its antigen is an element of its amino acids. [125] Those days, there are two kind of location strategy which every now and again utilized as a part of immunosensor which are optical and electrochemical. However optical discovery transduction technique has experienced poor affectability when combined with radioimmunoassay, the short half-existence of radioactive specialists, worries of wellbeing perils, and transfer issues. Electrochemical identification beats issues related with different methods of discovery of immunoassays and immunosensors. Conversely, electrochemical immunoassays and immunosensors empower quick, straightforward, and practical discovery that is free of these issues. [125] Be that as it may, late progress in science and innovation has made an optical transduction technique another way towards exceptionally advanced mechanized instrument. Consequently, Optical and electrochemical location technique are increasing common significance for advancement of immunosensor. [126-128] As indicated by Ramirez et al. 2009, Immunosenors have been imagined to assume a vital part in the change of open wellbeing by giving applications to which fast discovery, high affectability, and specificity are vital, in zones, for example, clinical science, nourishment quality, and natural checking. The advancement of immunosensor for microbes and pathogen location has picked up an extraordinary arrangement of consideration because of its application in the purpose of care estimation (POC). [129-132] Some current examinations

appeared that immunosensor is generally investigated toward the location of disease/tumor. Since the customary diagnostics technique is poor in affectability, selectivity and tedious, immunosensor are turned out to be promising instruments for growth identification at beginning periods of disease. [133]

12. Biomimetic Biosensor

A biomimetic biosensor is a counterfeit or manufactured sensor that imitates the capacity of a common biosensor. These can incorporate aptasensors, where aptasensors utilize aptamers as the biocomponent. Aptamers were accounted for without precedent for the early 1990s where portrayed as fake nucleic corrosive ligands. Aptamers were therefore synthetically identified with nucleic corrosive tests, however carried on more like immunizer and indicating astounding adaptability contrasted with other bio-acknowledgment segments. [134-135] Aptamer are manufactured strands of nucleic corrosive that can be intended to perceive amino acids, oligosaccharides, peptides, and proteins. An aptamer has few points of interest over immune response based biosensor, for example, high restricting proficiency, maintaining a strategic distance from the utilization of creature (i.e. decreased moral issue), littler and less perplexing, and so forth. Nonetheless, normal test confronting aptasensors is that they innate the properties of nucleic acids, for example, auxiliary pleomorphic furthermore, synthetic effortlessness which decreased the test proficiency and furthermore increment its creation cost. Along these lines, some exertion has been coordinated towards portrayal and advancement of aptamer to defeat this impediment. Aptamer properties, for example, their high specificity, little size, change and immobilization adaptability, regenerability or conformational change instigated by the objective restricting have been effectively abused to upgrade an assortment of bio-

detecting groups. [134] Aptamer based biosensor has been generally utilized as a part of different application. As of late adequate advance has been made in biomimetic sensor and aptasensors for clinical application. [136] This including clinical diagnostics to recognize pathogen, infection and irresistible illness. [137-140]

13. Enzyme based Biosensor

Protein based biosensor are the most punctual biosensor among the various biosensor, these biosensor is first presented by Clark and Lyons in 1962 an amperometric chemical cathode for glucose sensor which utilize 'dissolvable' protein cathode. [141] Since the in the first place biosensor, protein based biosensor has confront an enormous development in utilization for different application till introduce. Proteins are exceptionally proficient biocatalysts, which can particularly perceive their substrates and to catalyze their change. These one of a kind properties make the chemicals capable instruments to create logical gadgets. [142] Protein based biosensors relate personally a biocatalyst-containing detecting layer with a transducer. Protein based biosensor working chief depends on synergist activity and restricting abilities for particular recognition (David et al. 2008). The chemical based biosensors were made of protein as bioreceptor which is particular to distinguish focused on analyte frame test grid. The bolt and key and prompted fit speculation can apply to clarify the component of catalyst activity which is very particular for this sort of biosensor. This particular synergist response of the catalyst furnishes these sorts of biosensor with the capacity to distinguish much lower limits than with typical restricting procedures. This high specificity of enzyme– substrate joint efforts and the commonly high turnover rates of biocatalysts are the origination of sensitive and specific impetus based biosensor contraptions. [143] In a perfect world catalyst synergist activity can be impact by a few factors, for example, the focus of the substrate, temperature, nearness

of focused and non-aggressive inhibitor and pH. [143] Basically the Michaelis-Menten condition can be utilized to additionally clarify the location farthest point of chemical based biosensor. [144] Glucose oxidase (GOD) and horseradish peroxidase (HRP) are the most generally utilized compound based biosensor that has been accounted for in writing. Be that as it may, some current investigations have demonstrated that catalyst based biosensor can be utilized to recognize cholesterol, nourishment security and natural checking, overwhelming metals and furthermore Pesticides. [145-149] In addition, a current investigations announced the utilized of catalyst reactant suggestion fusing with nucleic corrosive biosensor for DNA discovery. [150]

14. Cell based Biosensor

Cell based sensor are the sort of biosensor, which utilize living cell as the biospecific detecting component and depend on the capacity of living cell to recognize the intracellular and extra cellular microenvironment condition, physiological parameter and produces reaction through the collaboration amongst jolt and cell. [151] Microorganisms, for example, microscopic organisms also, parasites can be utilized as biosensors to recognize particular atoms or the in general "state" of the encompassing condition. [152] Also, proteins that are accessible in cells can in like manner be used as bioreceptor for the acknowledgment of specific analyte. [152] Basically, living cell based biosensor is a remarkable biosensor as opposed to other kind of biosensor that contains materials that separated from living things. These sorts of biosensor has utilized of advantages and disadvantages. The recognition furthest reaches of this biosensor is chiefly controlled by the normal natural conditions in which the cell can remain alive for long stretch where require the control the physical and synthetic parameter of condition. However the major confinement with cell based biosensor are the security of the cell, which relies upon different conditions, for example, the

disinfection, lifetime, biocompatibility and so on. Another issue that represents the accomplishment of a cell based sensor depends basically particle selectivity, in which cell based sensor has poor selectivity of microbial sensor due to the multireceptor conduct of the in place cells. [153] In spite of these entanglements the cell based biosensor still ideal among the scientist because of the favorable circumstances over the chemicals based biosensor. The cell based biosensor are less delicate to hindrance by solutes and are more tolerant of problematic pH furthermore, temperature esteems than chemical based biosensor, however they should not surpass the thin run if there should arise an occurrence of the cells kicking the bucket, a more extended lifetime can be normal than with the enzymatic sensors and they are significantly less expensive in light of the fact that dynamic cells don't should be confined. [154] Writing uncovered that, cell based sensor have turned into a rising devices for restorative diagnostics (i.e. for example, sickness location), ecological examination, sustenance quality control, synthetic pharmaceutical industry and medications recognition. [155-161]

CONCLUSIONS

Biosensors offer an exciting alternative to traditional methods that combines living part with physicochemical identifier part and allowing rapid "real-time" and multiple analyses for diagnosis and estimations. Biosensors technology has been developed encompassing a variety of applications and mainly emphasis on the development of sensing elements and transducers which are under current research.

REFERENCES

1. Thevenot DR, Toth K, Durst RA, Wilson GS. Electrochemical biosensors: recommended definitions and classification. *Pure Appl Chem*. 1999; 71:2333–2348.
2. Thevenot DR, Toth K, Durst RA, Wilson GS. Electrochemical biosensors:

- recommended definitions and classification. *Biosens Bioelectron*. 2001; 16:121–131.
3. Thevenot DR, Toth K, Durst RA, Wilson GS. Electrochemical biosensors: recommended definitions and classification. *Anal Lett*. 2001; 34:635–659.
4. Wang JX, Sun XW, Wei A, Lei Y, Cai XP, Li CM, et al. Zinc oxide nanocomb biosensor for glucose detection. *Applied Physics Letters* [Internet]. AIP Publishing; 2006 Jun 5; 88(23):233106.
5. Hinze S. Bibliographical cartography of an emerging interdisciplinary discipline: The case of bioelectronics. *Scientometrics* [Internet]. Springer Nature; 1994 Mar;29(3):353–76.
6. Sharma SK, Sehgal N, Kumar A. Biomolecules for development of biosensors and their applications. *Current Applied Physics* [Internet]. Elsevier BV; 2003 Apr;3(2-3):307–16.
7. Schaertel BJ, Firstenberg-EDEN R. Biosensors in the Food Industry: Present and Future. *Journal of Food Protection* [Internet]. International Association for Food Protection; 1988 Oct; 51(10):811–20.
8. Highlights. Lab on a Chip [Internet]. Royal Society of Chemistry (RSC); 2001;1(2):21N. Available from: <http://dx.doi.org/10.1039/b109714h>.
9. Burkle FM. Measures of Effectiveness in Large-scale Bioterrorism Events. *Prehospital and Disaster Medicine* [Internet]. Cambridge University Press (CUP); 2003 Sep;18(03):258–62.
10. Mascini M. Affinity electrochemical biosensors for pollution control. *Pure and Applied Chemistry* [Internet]. Walter de Gruyter GmbH; 2001 Jan 1;73(1):23–30.
11. Malhotra BD, Chaubey A. Biosensors for clinical diagnostics industry. *Sensors and Actuators B: Chemical* [Internet]. Elsevier BV; 2003 Jun;91(1-3):117–27.
12. Brouwers A, Bernard D, Langlois M. Clinical and Analytical Performance of the Accu-Chek Inform Point-of-Care Glucose Meter. *Point of Care: The Journal of Near-Patient Testing & Technology* [Internet]. Ovid Technologies (Wolters Kluwer Health); 2005 Mar; 4(1):36–40.
13. Anjum V, Pundir CS. Biosensors: Future analytical tools. *Sensors & Transducers Journal*; 2007 Feb; 76: 937-944.
14. Palmer AE, Qin Y, Park JG, McCombs JE. Design and application of genetically

- encoded biosensors. Trends in Biotechnology [Internet]. Elsevier BV; 2011 Mar;29(3):144–52.
15. Situ C, Mooney MH, Elliott CT, Buijs J. Advances in surface plasmon resonance biosensor technology towards high-throughput, food-safety analysis. TrAC Trends in Analytical Chemistry [Internet]. Elsevier BV; 2010 Dec;29(11):1305–15.
 16. Tamayo J, Humphris AD., Malloy A., Miles M. Chemical sensors and biosensors in liquid environment based on microcantilevers with amplified quality factor. Ultramicroscopy [Internet]. Elsevier BV; 2001 Jan;86(1-2):167–73.
 17. Edelstein R. The BARC biosensor applied to the detection of biological warfare agents. Biosensors and Bioelectronics [Internet]. Elsevier BV; 2000 Jan;14(10-11):805–13.
 18. Corcuera, D.R.J. and Cavalieri, P.R. Biosensors. Encyclopedia of Agricultural, Food, and Biological Engineering.(2003).
 19. Borgmann, S; Schulte, A; Neugebauer S and Schuhmann, W. Amperometric Biosensors. *Advances in Electrochemical Science and Engineering.* (2011).
 20. Wang J. Electrochemical Glucose Biosensors. Chemical Reviews [Internet]. American Chemical Society (ACS); 2008 Feb;108(2):814–25.
 21. Akyilmaz E, Yorganci E, Asav E. Do copper ions activate tyrosinase enzyme? A biosensor model for the solution. Bioelectrochemistry. 2010;78:155–160.
 22. Venugopal V. Biosensors in fish production and quality control. Biosens Bioelectron. 2002; 17:147–157.
 23. Diviès C. Remarques sur l'oxydation de l'éthanol par une electrode microbienne d'acetobacter zylinum. Ann Microbiol. 1975;126A:175–186.
 24. Rechnitz GA. Biochemical electrodes uses tissues slice. Chem Eng News. 1978;56:16–21.
 25. Jensen, J.N. and A.M. Dietrich, Chemical Species. Water Environment Research, 1994. 66(4): p. 279-291.
 26. Bakker E, Telting-Diaz M. Electrochemical Sensors. Analytical Chemistry [Internet]. American Chemical Society (ACS); 2002 Jun;74(12):2781–800.
 27. Bakker E. Electrochemical Sensors. Analytical Chemistry [Internet]. American Chemical Society (ACS); 2004 Jun;76(12): 3285–98.
 28. Bakker E, Diamond D, Lewenstam A, Pretsch E. Ion sensors: current limits and new trends. Analytica Chimica Acta [Internet]. Elsevier BV; 1999 Jun;393(1-3):11–8.
 29. Rogers, K.R. & Mascini, M., Biosensors for field analytical monitoring, *Field Analytical Chemistry & Technology*, 1998, 2(6), 317-331.
 30. Leatherbarrow RJ, Edwards PR. Analysis of molecular recognition using optical biosensors. Curr Opin Chem Biol. 1999; 3:544–547.
 31. Badley, R.A., et al., Optical Biosensors for Immunoassays: The Fluorescence Capillary-Fill Device [and Discussion]. Philosophical Transactions of the Royal Society of London. Series B, Biological Sciences, 1987. 316(1176):143-160
 32. Louie, A.S., Marenchic, I.G. & Whelan, R.H., A fieldable modular biosensor for use in detection of foodborne pathogens, *Field Analytical Chemistry & Technology*, 1998;2(6): 371-377.
 33. Muhammad-Tahir, Z. & Alocilja, E.C. A disposable biosensor for pathogen detection in fresh produce samples, *Biosystems Engineering*, 2004; 88(2):145-151.
 34. Muhammad-Tahir Z, Alocilja EC. A conductometric biosensor for biosecurity. Biosensors and Bioelectronics [Internet]. Elsevier BV; 2003 May;18(5-6):813–9.
 35. Jensen, J.N. and A.M. Dietrich, Chemical Species. Water Environment Research, 1994; 66(4): 279-291.
 36. Katzir S, editor. The beginnings of piezoelectricity. Boston studies in philosophy of science [Internet]. Springer Netherlands; 2006; Available from: <http://dx.doi.org/10.1007/978-1-4020-4670-4>
 37. Steinem C, Janshoff A, editors. Piezoelectric Sensors. Springer Series on Chemical Sensors and Biosensors [Internet]. Springer Berlin Heidelberg; 2007; Available from: <http://dx.doi.org/10.1007/b100347>
 38. Tichý J, Erhart J, Kittinger E, Přívratská J. Piezoelectric Properties. Fundamentals of Piezoelectric Sensorics [Internet]. Springer Berlin Heidelberg; 2010;69–100. Available from: http://dx.doi.org/10.1007/978-3-540-68427-5_5
 39. Nicu L, Guirardel M, Chambosse F, Rougerie P, Hinh S, Trevisiol E, et al. Resonating piezoelectric membranes for

- microelectromechanically based bioassay: detection of streptavidin–gold nanoparticles interaction with biotinylated DNA. *Sensors and Actuators B: Chemical* [Internet]. Elsevier BV; 2005 Sep;110(1):125–36.
40. Vaughan RD, Guilbault GG. Piezoelectric Immunosensors. *Springer Series on Chemical Sensors and Biosensors* [Internet]. Springer Berlin Heidelberg; 2007; 237–80. Available from: http://dx.doi.org/10.1007/978-3-540-36568-6_7
 41. Serra B, Gamella M, Reviejo AJ, Pingarrón JM. Lectin-modified piezoelectric biosensors for bacteria recognition and quantification. *Anal Bioanal Chem*. 2008;391: 1853–1860.
 42. Tothill IE. Biosensors for cancer markers diagnosis. *Semin Cell Dev Biol*.2009; 20: 55–62.
 43. Nirschl M, Blüher A, Erler C, Katzschner B, Vikholm-Lundin I, Auer S, Vörös J, Pompe W, Schreiter M, Mertig M. Film bulk acoustic resonators for DNA and protein detection and investigation of in vitro bacterial S-layer formation. *Sens Actuators A Phys*. 2009;156: 180–184.
 44. Skládal P, Riccardi CDS, Yamanaka H, Da Costa PI. Piezoelectric biosensors for real-time monitoring of hybridization and detection of hepatitis C virus. *J Virol Methods*.2004; 117: 145–151.
 45. Yao C, Zhu T, Tang J, Wu R, Chen Q, Chen M, Zhang B, Huang J, Fu W. Hybridization assay of hepatitis B virus by QCM peptide nucleic acid biosensor. *Biosens. Bioelectron*. 2008;23: 879–885.
 46. Chen H, Hu QY, Yue-Zheng, Jiang JH, Shen GL, Yu RQ. Construction of supported lipid membrane modified piezoelectric biosensor for sensitive assay of cholera toxin based on surface-agglutination of ganglioside-bearing liposomes. *Anal. Chim. Acta*.2010; 657: 204– 209.
 47. Chen SH, Wu VCH, Chuang YC, Lin CS. Using oligonucleotide-functionalized Au nanoparticles to rapidly detect foodborne pathogens on a piezoelectric biosensor. *J Microbiol Meth*. 2008;73: 7–17.
 48. Dell'Atti D, Tombelli S, Minunni M, Mascini M. Detection of clinically relevant point mutations by a novel piezoelectric biosensor. *Biosens Bioelectron*. 2006;21: 1876–1879.
 49. Glynn B, O'Connor L. Nucleic Acid Diagnostic Biosensors. In Zourob M (ed): *Recognition receptors in biosensors*. Springer, New York 2010; 343-364.
 50. Leca-Bouvier BD, Blum LJ. Enzyme for Biosensing Applications. In Zourob M (ed): *Recognition receptors in biosensors*. Springer, New York 2010; 177-220.
 51. Kim JM, Chang SM, Muramatsu H, Isao K. The principles and applications of nano diagnosis system for a nano-biosensor. *Korean J Chem Eng*.2011; 28: 987–1008.
 52. Liu A, Wang K, Weng S, Lei Y, Lin L, Chen W, Lin X, Chen. (2012). Development of electrochemical DNA biosensors. *Trends. Anal. Chem*. 2012;37: 101–111.
 53. Parkinson G, Pejcic B. Using Biosensors to Detect Emerging Infectious Diseases, prepared for The Australian Biosecurity Cooperative Research Centre.2005;1-80.
 54. Cagnin S, Caraballo M, Guiducci C, Martini P, Ross M, Santaana M, Danley D, West T Lanfranchi G. Overview of electrochemical DNA biosensors: new approaches to detect the expression of life. *Sensor*.2009; 9: 3122–3148.
 55. Teles F, Fonseca L. Trends in DNA biosensors. *Talanta*. 2008;77: 606–623.
 56. Brett AMO. DNA based biosensors. In Gorton L (ed): *Comprehensive Analytical Chemistry XLIV: Biosensors and Modern Biospecific Analytical Techniques*. Elsevier, Amsterdam 2005; 179-208.
 57. Chua A, Yean CY, Ravichandran M, Lim B, Lalitha P. A rapid DNA biosensor for the molecular diagnosis of infectious disease. *Biosens Bioelectron*.2011; 26: 3825–3831.
 58. Liu CC. Electrochemical Based Biosensors. *Biosensors*. 2012;2: 269–272.
 59. Thuy NT, Tam PD, Tuan MA, Le AT, Tam LT, Thu VV, Hieu NV, Chien ND. Detection of pathogenic microorganisms using biosensor based on multi-walled carbon nanotubes dispersed in DNA solution. *Curr Appl Phys*. 2012;12: 1553–1560.
 60. Arora K, Prabhakar N, Chand S, Malhotra BS Immobilization of single stranded DNA probe on polypyrrole-polyvinyl sulfonate for application to DNA hybridization biosensor. *Sensors Actuators B*.2007; 126:655-663.
 61. Yau HCM, Chan HL, Yang M Electrochemical properties of DNA intercalating doxorubicin and methylene

- blue on n-hexadecylmercaptandoped 5'-thiol-labeled DNA-modified gold electrodes. *Biosensors Bioelectronics*.2003; 18: 873-879.
62. Gu J, Lu X, Ju H. DNA sensor for recognition of native yeast DNA sequence with methylene blue as an electrochemical hybridization indicator. *Electro Analysis*.2002; 14: 949-953.
63. Reddy RRR, Chadha A, Bhattacharya E. Porous silicon based potentiometric triglyceride biosensor. *Biosens Bioelectron*.2001; 16: 313-317.
64. Kara P, Kerman K, Ozkan D, Meric B, Erdem A, Ozkan Z, et al. Electrochemical genosensor for the detection of interaction between methylene blue and DNA. *Electrochemistry Communications* [Internet]. Elsevier BV; 2002 Sep;4(9):705-9.
65. Erdem A, Kerman K, Meric B, Ozsoz M. Methylene blue as a novel electrochemical hybridization indicator. *Electro analysis* .2001;13: 219-223.
66. Arora K, Malhotra BD. Applications of conducting polymer based DNA biosensors. *Applied Physics in the 21st century*. 2008; 1-20.
67. Kerman K, Kobayashi M, Tamiya E. Recent trends in electrochemical DNA biosensor technology. *Meas Sci Technol*.2004; 15: 1-11.
68. Lucarelli F, Tombelli S, Minunni M, Marrazza G, Mascini M. Electrochemical and piezoelectric DNA bio-sensor for hybridization detection. *Anal Chim Acta*.2008; 609: 139-159.
69. Cheng G, Zhao J, Tu Y, He P, Fang Y. A sensitive DNA electrochemical biosensor based on magnetite with a glassy carbon electrode modified by multi-walled carbon nanotubes in poly-pyrrole. *Anal Chim Acta*.2005; 533: 11-16.
70. Mascini M, Palchetti I, Murazza G. DNA electrochemical biosensor. *Fresenius J Anal Chem*.2001; 369: 15-22.
71. Chrisey LA, Lee GU, Ferrall CEO. Covalent attachment of synthetic DNA to self assembled monolayer films. *Nucl Acids Res* .1996; 24:3031-3039.
72. Marks RS, Cullen DC, Karube I, Lowe CR, Weetall HH, editors. *Handbook of Biosensors and Biochips*. John Wiley & Sons, Ltd; 2008 Mar 15; Available from: <http://dx.doi.org/10.1002/9780470061565>
73. Wang J. Survey and Summary: From DNA biosensors to gene chips. *Nucleic Acids Research* [Internet]. Oxford University Press (OUP); 2000 Aug 15;28(16):3011-3016.
74. Marrazza G, Chinella I, Mascini M. Disposable DNA electrochemical sensor for hybridization detection. *Biosens Bioelectron*.1999; 14: 43-51.
75. Arora K, Prabhakar N, Chand S, Malhotra BD (2007) Escherichia coli genosensor based on polyaniline. *Anal Chem*. 2007;79: 6152-6158.
76. Pingang H, Ying X, Yuzhi F. Application of carbon nanotubes in electrochemical DNA biosensor(Review). *Microchim Acta*.2006; 152: 175-186.
77. Arora K, Prabhakar N, Chand S, Malhotra BD. Ultrasensitive DNA hybridization biosensor based on polyaniline. *Biosens Bioelectronics*.2007; 23:613-620.
78. Yang W, Ozsoz M, Hibbert DB, Gooding JJ. Evidence for the direct interaction between methylene blue and guanine bases using DNA modified carbon paste electrodes, *Electro analysis*.2002; 14: 1299-1302.
79. Campbell CN, Gal D, Cristler N, Banditrat C, Heller A. Enzyme amplified amperometric sandwich test for RNA and DNA. *Anal Chem*.2002;74: 158-162.
80. Prabhakar N, Arora K, Singh SP, Singh H, Malhotra BD. DNA entrapped polypyrrole-polyvinyl sulfonate film for application to electrochemical biosensor. *Anal Biochem* .2007;366: 71-79.
81. Yanmashita K, Takagi M, Kondo H, Takenaka S. Electrochemical detection of nucleic base mismatches with ferrocenyl naphthalene diimide. *Anal Biochem*.2002; 306: 188-196.
82. Tombelli S, Mascini M, Braccini L, Anichini M, Turner APF. Coupling of a DNA piezoelectric biosensor and polymerase chain reaction to detect apolipoprotein- E polymorphisms. *Biosens Bioelectron*. 2000;15:363-370.
83. Mol NJD, Fischer MJE. Surface Plasmon Resonance: A General Introduction. In Mol NJD Fischer MJE (ed): *Surface Plasmon Resonance Methods and Protocols*. Springer, New York. 2010;pp. 1-14.
84. Endo T, Yamamura S, Nagatani N, Morita Y, Takamura Y, Tamiya E. Localized surface plasmon resonance based optical

- biosensor using surface modified nanoparticle layer for label-free monitoring of antigen-antibody reaction. *Sci Technol Adv Mat.*2005; 6: 491-500.
85. Mannelli I, Courtois V, Lecaruyer P, Roger G, Millot MC, Goossens M, Canva M. Surface plasmon resonance imaging (SPRI) system and real-time monitoring of DNA biochip for human genetic mutation diagnosis of DNA amplified samples. *Sens Actuators B Chem.*2006; 119: 583–591.
 86. Nguyen B, Tanius FA, Wilson WD. Biosensor-surface plasmon resonance: quantitative analysis of small molecule-nucleic acid interactions. *Methods.* 2007;42: 150–161.
 87. Nakamura H, Mogi Y, Akimoto T, Naemura K, Kato T, Yano K, Karube I. An enzyme chromogenic surface plasmon resonance biosensor probe for hydrogen peroxide determination using a modified Trinder's reagent. *Biosens Bioelectron.* 2008;24: 455–460.
 88. Park TJ, Hyun MS, Lee HJ, Lee SY, Ko S. A self-assembled fusion protein-based surface plasmon resonance biosensor for rapid diagnosis of severe acute respiratory syndrome. *Talanta.* 2009; 79: 295–301.
 89. Fan X, White IM, Shopova SI, Zhu H, Suter JD, Sun Y. Sensitive optical biosensors for unlabeled targets: a review. *Anal Chim Acta.*2008; 620: 8–26.
 90. Endo T, Yamamura S, Kerman K, Tamiya E. Label-free cell-based assay using localized surface plasmon resonance biosensor. *Anal Chim Acta.*2008; 614: 182–189.
 91. Homola J, Yee SS, Myszka D. Surface plasmon resonance biosensors. In Ligler F, Taitt (ed): *Optical biosensors: today and tomorrow.* Elsevier, Amsterdam. 2008;185-242.
 92. Merwe PAVD. Surface plasmon resonance. In Harding S, Chowdhry PZ (ed): *Protein-Ligand interactions: hydrodynamics and calorimetry.* Oxford University Press, Bath.2001; 137-170.
 93. Mitchell J, Wu Y. Surface Plasmon Resonance Biosensors for Highly Sensitive Detection of Small Biomolecules. In Serra PA (ed): *Biosensors.* InTech, Rijeka 2010;151-168.
 94. Shankaran D, Gobi K, Miura N. Recent advancements in surface plasmon resonance immunosensors for detection of small molecules of biomedical, food and environmental interest. *Sens Actuators B Chem.*2007; 121: 158–177.
 95. Skottrup PD, Nicolaisen M, Justesen AF. Towards on-site pathogen detection using antibody-based sensors. *Biosens Bioelectron.*2008; 24: 339–348.
 96. Caygill RL, Blair GE, Millner PA. A review on viral biosensors to detect human pathogens. *Anal Chim Acta.*2010; 681: 8–15.
 97. Li D. Chemiluminescence. In Li D (ed): *Encyclopedia of Microfluidics and Nanofluidics.* Springer, New York. 2008; pp. 260.
 98. Atias D, Liebes Y, Chalifa-Caspi V, Bremand L, Lobel L, Marks RS, Dussart P. Chemiluminescent optical fiber immunosensor for the detection of IgM antibody to dengue virus in humans. *Sens Actuators B Chem.*2009; 140: 206–215.
 99. Ding C, Zhong H, Zhang S. Ultrasensitive flow injection chemiluminescence detection of DNA hybridization using nanoCuS tags. *Biosens Bioelectron.* 2008; 23: 1314–1318.
 100. Holford TRJ, Davis F, Higson SPJ. Recent trends in antibody based sensors. *Biosens Bioelectron.*2012; 34: 12–24.
 101. Guo Z, Yang F, Zhang L, Zheng X. Electrogenated chemiluminescence energy transfer for the label-free detection of DNA. *Sens Actuators B Chem.* 2013;177: 316–321.
 102. Kricka L. Clinical applications of chemiluminescence. *Anal Chim Acta.* 2003;500: 279–286.
 103. Gámiz-gracia L, García-campa AM, Huertas-pérez JF, Lara FJ. Chemiluminescence detection in liquid chromatography: Applications to clinical , pharmaceutical ,environmental and food analysis - A review. *Anal Chim Acta.* 2009; 640: 7–28.
 104. Lara FJ, García-campa AM, Aaron J. Analytical applications of photoinduced chemiluminescence in flow systems - A review. *Anal Chim Acta.* 2010;679: 17–30.
 105. Liu M, Lin Z, Lin J. A review on applications of chemiluminescence detection in food analysis. *Anal Chim Acta.*2010; 670: 1–10.
 106. Parkinson G, Pejčić B. Using Biosensors to Detect Emerging Infectious Diseases, prepared for The Australian

- Biosecurity Cooperative Research Centre:2005 pp.1-80.
107. Zhang, Z., Zhang, S., & Zhang, X.: Recent developments and applications of chemiluminescence sensors. *Anal Chim Acta.* 2005;541: 37–46.
 108. Mathews ST, Plaisance EP, Kim T. Imaging Systems for Westerns: Chemiluminescence vs. Infrared Detection. In Kurien BT, Scofield RH (ed): *Methods in Molecular Biology, Protein Blotting and Detection.* Humana Press, Totowa 2009, pp. 499-513.
 109. Dodeigne C, Thunus L, Lejeune R. Chemiluminescence as diagnostic tool. A review. *Talanta.* 2000;51: 415–439.
 110. Daly CJ, McGrath JC. Fluorescent ligands, antibodies, and proteins for the study of receptors. *Pharmacol Therapeut.* 2003; 100: 101–118.
 111. Biran I, Yu X, Walt DR. Optrode-based fiber optic biosensors (bio-optrode). In Ligler F, Taitt C (ed): *Optical biosensors: today and tomorrow.* Elsevier, Amsterdam. 2008; 3- 82.
 112. Ramanathan A, Pape L, Schwartz DC. High-density polymerase-mediated incorporation of fluorochrome-labeled nucleotides. *Anal biochem.*2005; 337: 1–11.
 113. Berdat D, Marin A, Herrera F, Gijns MAM. DNA biosensor using fluorescence microscopy and impedance spectroscopy. *Sens Actuators B Chem.* 2006; 118: 53–59.
 114. Schultz E, Galland R, Du Bouëtiez D, Flahaut T, Planat-Chrétien A, Lesbre F, Hoang A, Volland H, Perraut F. A novel fluorescence-based array biosensor: principle and application to DNA hybridization assays. *Biosens Bioelectron.*2008; 23: 987–994.
 115. Hong B, Kang KA. Biocompatible, nanogold-particle fluorescence enhancer for fluorophore mediated, optical immunosensor. *Biosens Bioelectron.*2006; 21: 1333–1338.
 116. Moghissi K, Stringer MR, Dixon K. Fluorescence photodiagnosis in clinical practice *Photodiagn Photodyn.*2008; 5: 235–237.
 117. Grubor NM, Shinar R, Jankowiak R, Porter MD, Small GJ. Novel biosensor chip for simultaneous detection of DNA-carcinogen adducts with low-temperature fluorescence. *Biosens Bioelectron.*2004; 19: 547–556.
 118. Thompson RB. Fluorescence lifetime biosensing: Entering the mainstream. In Ligler F, Taitt C (ed): *Optical biosensors: today and tomorrow.* Elsevier, Amsterdam 2008, 287-316.
 119. Abdulhalim I, Zourob M, Lakhtakia A. Overview of Optical Biosensing Techniques. In Marks RS, Lowe CR, Cullen DC, Weetall HH, Karube I (ed): *Handbook of Biosensors and Biochips.* Wiley, Weinheim.2007.
 120. Donahue AC, Albitar M. Antibodies in Biosensing. In Zourob M (ed): *Recognition receptors in biosensors.* Springer, New York.2010; 221-248.
 121. Conroy PJ, Hearty S, Leonard P, O’Kennedy RJ. Antibody production, design and use for biosensor-based applications. *Semin Cell Dev Biol.*2009; 20: 10–26.
 122. Orazio PD. Biosensors in clinical chemistry - 2011 update. *Clin Chim Acta.*2011; 412: 1749- 1761.
 123. Wood EJ. Lecture notes: Immunology (5th edition). *Biochemistry and Molecular Biology Education [Internet].* Wiley-Blackwell; 2006 May;34(3):235a–236.
 124. Pohanka M. Monoclonal and polyclonal antibodies production-preparation of potent biorecognition element. *J Appl Biomed.* 2009; 7: 115–121.
 125. Fowler JM, Wong DKY, Halsall HB, Heineman WR. Recent developments in electrochemical immunoassays and immunosensors. In Zhang X, Ju H, Wang J (ed): *Electrochemical Sensors, Biosensors and Their Biomedical Applications.* Elsevier, San Diego; 2008; 115-140.
 126. Shankaran D, Gobi K, Miura N. Recent advancements in surface plasmon resonance immunosensors for detection of small molecules of biomedical, food and environmental interest. *Sens Actuators B Chem.*2007; 121: 158–177.
 127. Bhatta D, Stadden E, Hashem E, Sparrow IJG, Emmerson GD. Label-free monitoring of antibody-antigen interactions using optical microchip biosensors. *J Immunol Method.* 2010b;362: 121–126.
 128. Bhatta D, Stadden E, Hashem E, Sparrow IJG, Emmerson GD. Multi-purpose optical biosensors for real-time detection of bacteria, viruses and toxins. *Sens Actuators B Chem.*2010a;149: 233–238.

129. Skottrup PD, Nicolaisen M, Justesen AF. Towards on-site pathogen detection using antibody-based sensors. *Biosens Bioelectron.*2008;24: 339–348.
130. Barton AC, Collyer SD, Davis F, Garifallou GZ, Tsekenis G, Tully E, O’Kennedy R, Gibson T, Millner PA, Higson SPJ. Labelless AC impedimetric antibody-based sensors with pgml(-1) sensitivities for point-of-care biomedical applications. *Biosens Bioelectron.*2009; 24: 1090–1095.
131. Braiek M, Rokbani KB, Chrouda A, Bakhrouf BA, Maaref A, Jaffrezic-Renault N. An Electrochemical Immunosensor for Detection of Staphylococcus aureus Bacteria Based on Immobilization of Antibodies on Self-Assembled Monolayers-Functionalized Gold Electrode. *Biosensors.*2012; 2: 417–426.
132. Holford TRJ, Davis F, Higson SPJ. Recent trends in antibody based sensors. *Biosens Bioelectron.*2012; 34: 12–24.
133. Ushaa SM, Nagar NI, Rao GM. Design and analysis of nanowire sensor array for prostate cancer detection. *Int J Nano Biomater.*2011; 3: 239–255.
134. Schneider HJ, Lim S, Strongin RM. Biomimetic Synthetic Receptors as Molecula Recognition Elements. In Zourob M (ed): Recognition receptors in biosensors. Springer, New York .2010;777-818.
135. Brys E, Tombelli S, Minunni M, Mascini M, Turner APF. Approaches to Allergy Detection Using Aptasensors. In Knopf GK, Bassi AS (ed): Smart Biosensor Technology. CRC Press, Boca Raton .2007; 539-566.
136. Vallet-Regí M, Arcos DA. Biomimetic Nanoceramics in Clinical Use: From Materials to Applications. RSC publishing, Cambridge.2008.
137. Strehlitz B, Nikolaus N, Stoltenburg R. Protein Detection with Aptamer Biosensors. *Sensors.* 2008;8: 4296–4307.
138. Torres-Chavolla E, Alocilja EC. Aptasensors for detection of microbial and viral pathogens *Biosens Bioelectron.*2009; 24: 3175–3182.
139. Wang YX, Ye ZZ, Si CY, Ying YB. Application of Aptamer Based Biosensors for Detection of Pathogenic Microorganisms. *Chinese J Anal Chem.*2012; 40: 634–642.
140. Weng CH, Huang CJ., Lee GB. Screening of Aptamers on Microfluidic Systems for Clinical Applications. *Sensors.*2012; 12: 9514–9529.
141. Shantilatha R, Varma S, Mitra CK. Designing a simple biosensor. In Biosensors: Perspectives. In Malhotra BD, Turner APF (ed): Advances In Biosensors: Perspectives In Biosensor. JAI Press, Stamford. 2003; 1-36.
142. Leca-Bouvier BD, Blum LJ. Enzyme for Biosensing Applications. In Zourob M (ed): Recognition receptors in biosensors. Springer, New York .2010; 177-220.
143. Cass T. Enzymology. In Marks RS, Lowe CR, Cullen DC, Weetall HH, Karube I (ed): Handbook of Biosensors and Biochips. Wiley, Weinheim. 2007.
144. Parkinson G, Pejicic B. Using Biosensors to Detect Emerging Infectious Diseases, prepared for The Australian Biosecurity Cooperative Research Centre.2005; pp.1-80.
145. Amine A, Mohammadi H, Bourais I, Palleschi G. Enzyme inhibition-based biosensors for food safety and environmental monitoring. *Biosens Bioelectron.* 2006;21: 1405–1423.
146. Zapp E, Brondani D, Vieira IC, Scheeren CW, Dupont J, Barbosa AMJ, Ferreira VS. Biomonitoring of methomyl pesticide by laccase inhibition on sensor containing platinum nanoparticles in ionic liquid phase supported in montmorillonite. *Sens Actuators B Chem.* 2011;155: 331–339.
147. Nomngongo PN, Catherine Ngila J, Msagati TAM, Gumbi BP, Iwuoha EI. Determination of selected persistent organic pollutants in wastewater from landfill leachates, using an amperometric biosensor. *Phys Chem Earth Pt A/B/C.* 2012;50-52: 252–261.
148. Soldatkin OO, Kucherenko IS, Pyeshkova VM, Kukla AL, Jaffrezic-Renault N, El’skaya AV, Dzyadevych SV, Soldatkin AP. Novel conductometric biosensor based on three enzyme system for selective determination of heavy metal ions. *Bioelectrochemistry.* 2012;83: 25–30.
149. Ju H, Kandimalla VB. Biosensors for pesticides. In Zhang X, Ju H, Wang J (ed) Electrochemical Sensors, Biosensors and Their Biomedical Applications. Elsevier, San Diego. 2008, pp. 32-50.

150. He Y, Zhang S, Zhang X, Baloda M, Gurung AS, Xu H, Zhang X, Liu G. Ultrasensitive nucleic acid biosensor based on enzyme-gold nanoparticle dual label and lateral flow strip biosensor. *Biosens Bioelectron.*2011; 26: 2018–2024.
151. Wang P, Liu Q. *Cell-based biosensors: Principles and Applications.* Artech House, Norwood 2010.
152. Acha V, Andrews T, Huang Q, Sardar DK, Hornsby PJ. *Tissue-Based Biosensors.* In Zourob M (ed): *Recognition receptors in biosensors.* Springer, New York .2010; pp. 365- 382.
153. Belkin S, Gu MB. *Whole Cell Sensing Systems I: Reporter Cells and Devices.* Springer, Hiedelberg.2010.
154. Struss AK, Pasini P, Daunert S. *Biosensing Systems Based on Genetically Engineered Whole Cells.* In Zourob M (ed): *Recognition receptors in biosensors.* Springer, New York 2010, pp. 565-598.
155. Veisheh M, Veisheh O, Martin MC, Bertozzi C, Zhang M. Single-cell-based sensors and synchrotron FTIR spectroscopy: a hybrid system towards bacterial detection. *Biosens Bioelectron.*2007; 23: 253–260.
156. Banerjee P, Bhunia AK. Mammalian cell-based biosensors for pathogens and toxins. *Trends Biotechnol.* 2009;27: 179–88.
157. Banerjee P, Bhunia AK. Cell-based biosensor for rapid screening of pathogens and toxins. *Biosens Bioelectron.*2010; 26: 99–106.
158. Wang TH, Hui GH, Deng SP. A novel sweet taste cell-based sensor. *Biosens Bioelectron.* 2010;26: 929–934.
159. Melamed S, Elad T, Belkin S. Microbial sensor cell arrays. *Curr Opin Biotechnol.*2012;23: 2–8.
160. Bohr U, Stütz E, Fuchs K, Fleischer M, Schöning MJ, Wagner P. Monitoring of irritant gas using a whole-cell-based sensor system. *Sens Actuators B Chem.* 2012;175: 208–217.
161. Liu Q, Hu N, Zhang F, Wang H, Ye W, Wang P. Neurosecretory cell-based biosensor: monitoring secretion of adrenal chromaffin cells by local extracellular acidification using light-addressable potentiometric sensor. *Biosens Bioelectron.*2012; 35: 421–424.

How to cite this article: Kaur H, Bhosale A, Shrivastav S. Biosensors: classification, fundamental characterization and new trends: a review. *Int J Health Sci Res.* 2018; 8(6):315-333.
